

### Bid Document

Bid Details	
Bid End Date/Time	10-02-2020 09:00:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	70 (Days)
Ministry/State Name	Ministry Of Science And Technology
Department Name	Department Of Scientific And Industrial Research (dsir)
Organisation Name	N/a
Office Name	Csir - Central Salt & Marine Chemical Research Institute
Total Quantity	1
Item Category	two pan laboratory balance analyticalchemicalphysical
MSE Exemption For Years Of Experience And Turnover	Yes
Startup Exemption For Years Of Experience And Turnover	Yes
Document required from seller	Certificate (Requested in ATC), OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No

### EMD Detail

Required	No
----------	----

### ePBG Detail

Required	No
----------	----

### Splitting

Bid splitting not applied.

## Two Pan Laboratory Balance Analyticalchemicalphysical ( 1 pieces )

### Technical Specifications

[\\* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	<b>Conformity to Indian Standard, IS:9440 latest</b>	Yes	Yes
	<b>BIS Marked</b>	No	No
	CML No if BIS marked yes otherwise put NA	-	*
	Two Pan laboratory Balance must be in Compliance to ISO GLP Standards	Yes	*
	<b>Pan shall be as per Standard of Weights and Measures General Rules,1987</b>	Yes	Yes
Performance Parameters	<b>Type of Laboratory Balance</b>	Analytical	Analytical
	<b>Sensitivity (mg)</b>	0.1 mg	0.1 mg
	Type of Calibration	Internal	*
	<b>Maximum Capacity of weighing (grams)</b>	220 g	220 g
	Response Time in seconds	3	*
	<b>Shape of PAN</b>	Circular	Circular
	Power Supply	Single Phase	*
	<b>Warranty (Year)</b>	1	1
Dimensional Parameters	Material for Case	-	*
	Minimum Overall Diameter of Pan (mm)	90	*
	Material of Pointer, Levelling Arrangement, Pointer Scale , Rider Arrangement , Pointer Adjusting Screws and Stirrups must be as per Indian Standards 9440 latest	Yes	*
Additional Parameters	Tare Facility	Yes	*
	Availability of Permanent Shock	Yes	*

	Absorption facility		
	Overload Protection	Yes	*
	Marking of Left and Right to PAN	Yes	*
	Packing of Balance should be in convenient wooden case suitably padded and fitted with proper supports at various points	Yes	*
	<b>Printer and PC Connectivity</b>	Yes	Yes
	if Printer and PC connectivity is Yes mode of Connectivity (USB or RS232) otherwise put NA	-	*
	<b>Piece counting Features</b>	Yes	Yes
	<b>Totaling</b>	Yes	Yes
	<b>Check Weighing</b>	Yes	Yes
	Operating temperature Range (-10 Degree to 50 Degree Celcius)	Yes	*
Reports	Model Approval certificate for each model from director Legal Metrology	Yes	*
	Model Approval certificate No and date for each model	-	*
	Availability of Type Test reports From Central/Govt/ NABL/ILAC Accredited lab for weighing scale as per 9440 latest	Yes	*
	Test Report No	-	*
	Test Report Date	-	*
	Name of Lab	-	*
	Complete Address of Lab	-	*
	Calibration from concerned weight and measurement controller to be obtained at the time of supplies	Yes	*

	Model approval certificate, Type test report and calibration certificate to be furnished if asked by buyer at the time of supplies	Yes	*
--	--	-----	---

\* Specifications highlighted in bold are the Golden Parameters.

\* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

### Additional Specification Documents

### Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Patel Rakeshkumar Mavjibhai	364001, CSIR-Central Salt & Marine Chemicals Research Institute Gijubhai Badheka Marg Bhavnagar 364001	1	30

### Bid Specific Additional Terms and Conditions

1. Experience Criteria: The Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for 2 years before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the primary product having highest value should meet this criterion.
2. The buyer organization is an institution eligible for concessional rates of GST as notified by the Government of India. The goods for which bids have been invited fall under classification of GST concession and the conditions for eligibility of concession are met by the institution. A certificate to this effect will be issued by Buyer to the Seller after award of the Contract. Sellers are requested to submit their bids after accounting for the Concessional rate of GST. Applicable Concessional rate of GST : 5%  
Notification No. and date : 45/2017 and 47/2017 dated 14/11/2017
3. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods
4. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---